



Standard Specification for Serological Pipet, Disposable Plastic¹

This standard is issued under the fixed designation E 934; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers disposable plastic serological pipets, calibrated “to deliver” when measuring volumes of liquids.

1.1.1 *Any institution or individual who reuses a disposable pipet must bear full responsibility for its safety and effectiveness.*

2. Referenced Documents

2.1 ASTM Standards:²

D 703 Specification for Polystyrene Molding and Extrusion Materials³

E 542 Practice for Calibration of Laboratory Volumetric Apparatus

E 920 Specification for Commercially Packaged Laboratory Apparatus

E 921 Specification for Export Packaged Laboratory Apparatus

E 1133 Practice for Performance Testing of Packaged Laboratory Apparatus for United States Government Procurements

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *accuracy*—the expected distribution of mean volumes around the stated volume.

3.1.2 *coefficient of variation*—the expected distribution of individual volumes around the mean volume.

3.1.3 *disposable pipet*—such pipets will only be expected to provide their specified performance during their original use or operation.

4. Material and Manufacturer

4.1 The pipets made to this specification shall be fabricated from crystal grade, uncolored polystyrene, or regrind of same, in accordance with Specification D 703.

5. Design

5.1 *Shape*—0.5, 1.0, and 2.0-cm³ pipets shall be straight and of one-piece construction. Any cross section of a pipet taken in a plane perpendicular to the longitudinal axis shall be circular.

5.1.1 Pipets of 5.0, 10.0, 25.0, and 50.0 cm³ shall be straight and may consist of one, two, or three components, the extruded pipet barrel, the pulled or injection molded tip, and the plain, pulled, or injection molded top end. Any cross section of a pipet taken in a plane perpendicular to the longitudinal axis shall be circular.

5.2 *Delivery Tips*—Delivery tips shall be made with a gradual taper of 10 to 40 mm. The tip end shall be reasonably perpendicular to the longitudinal axis of the pipet, and shall be free of internal flash.

5.3 *Top End*—The 5, 10, 25.0, and 50.0-cm³ sizes shall have a top end with an inside diameter of 2 to 6.5 mm for a minimum distance of 20 mm from the open end, and shall have an overall length of 24 to 28 mm. On all sizes, the top end shall be suitable for plugging with filtering material. The top may be flat or chamfered and must be without sharp outer edges. All top ends shall be reasonably perpendicular to the longitudinal axis of the pipet. The O.D. of top end should range from 7 to 9 mm.

5.4 *Dimensions and Outflow Times*—The limiting dimensions and outflow times shall be as shown in Table 1. Outflow times shall be determined by means of a stopwatch on unplugged pipets when using distilled water at $25 \pm 5^\circ\text{C}$.

6. Markings

6.1 *Graduation Markings*—Graduation lines shall be between 0.2 mm and 0.5 mm in width, and in a plane perpendicular to the longitudinal axis of the pipet parallel to each other. A main graduation line shall extend at least $\frac{3}{5}$ of the way around the pipet. The values of all main graduation lines shall be in Arabic numbers directly above the lines referenced. Intermediate graduation lines shall extend at least $\frac{1}{5}$ of the way

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Apparatus.

Current edition approved May 1, 2004. Published May 2004. Originally approved in 1983. Last previous edition approved in 1998 as E 934 – 94 (1998).

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Withdrawn.

TABLE 1 Dimensions and Outflow Times

Capacity, cm ³	Least Value	Graduation cm ³ (mL)		Outside Diameter Graduated Portion, mm	Outflow, s		Wall Thickness, min, mm
		Main No.	Graduation 0 to		min	max	
0.5	0.01	0.1	0.4	4.5 to 4.8	0.5	3.0	0.8
1.0	0.01	0.1	0.9	4.5 to 4.8	0.5	3.0	0.8
1.0	0.10	0.1	0.9	4.5 to 4.8	0.5	3.0	0.8
2.0	0.01	0.1	1.8	5.5 to 6.7	0.5	4.0	0.9
2.0	0.10	0.2	1.8	5.5 to 6.7	0.5	4.0	0.9
5.0	0.10	1	4.0	7.6 to 8.4	0.5	7.0	0.9
10.0	0.10	1	9.0	9.5 to 11.7	0.5	8.0	0.9
25.0	0.20 or 0.25	1 or 2	23.0	13.0 to 15.0	1.0	10.0	0.7
50.0	0.25	1 or 2	47.0	17.0 to 19.0	2.0	12.0	0.8

around the pipet, and least graduation lines shall extend at least 1/8 of the way around the pipet. The zero graduation line shall be at least 90 mm below the top of the pipet on all sides.

6.2 *Identification Markings*—Each pipet shall be marked with the manufacturer’s or vendor’s name or trademark, its capacity and subdivision, the symbols “TD” (“to deliver”) and “20°C”, and either a wide band or two narrow bands on the top end to signify that the last drop must be blown out to achieve full delivery.

7. Capacity

7.1 The pipet capacity, expressed as cm³ (or mL), and the expected deviation from the stated capacity, shall be stated on the product package or published data. The expected deviation from stated capacity is expressed as accuracy and coefficient of variation. Pipets shall be calibrated for delivery at 20°C.

7.1.1 The accuracy shall be determined as specified in 7.2 and shall be within the limits given in Table 2.

7.1.2 The coefficient of variation shall be determined as specified in 7.2 and shall be within the limits given in Table 2.

7.2 *Determination of Accuracy and Coefficient of Variation*—A minimum number of 30 pipets in accordance with Practice E 542, taken at random, shall be tested gravimetrically, using distilled water. Distilled water at 20 ± 1°C is drawn into the pipet and the volume adjusted as in the case of normal use. This is discharged into a tared weighing bottle and

the mass, *m*, is determined using a balance capable of weighing to ±0.0001 g. The volume, *v_x*, is computed from the relation:

$$v_x = \frac{m}{0.998} \quad (1)$$

7.2.1 Accuracy (A)

A = ($\bar{x} - v_1$)/*v*₁ × 100, in percent,
 \bar{x} = mean of sample measurements, and
*v*₁ = stated capacity or stated volume of pipet.

7.2.2 Coefficient of Variation (cv)

cv = 100*S*/ \bar{x} , in percent
S = standard deviation

$$= \sqrt{\frac{\sum(x - \bar{x})^2}{n - 1}}$$

x = individual sample measurements,
 \bar{x} = mean of sample measurements, and
n = number of pipets measured.

7.2.3 The use of volumetric standards developed from and traceable to gravimetric standards is permissible if said volumetric standards were calibrated using weighings and corrections as recommended in Practice E 542.

TABLE 2 Accuracy and Coefficient of Variation of Stated Capacity

Capacity cm ³	Accuracy + % ^A	Coefficient of Variation Equal to or less than 1.5 %
0.5	3	1.5
1.0	3	1.5
2.0	3	1.5
5.0	3	1.5
10.0	3	1.5
25.0	3	1.5
50.0	3	1.5

^A Accuracy and coefficient of variation are expressed as a percentage of the stated capacity.

8. Workmanship, Finish, and Appearance


8.1 Pipets shall be free from defects that may impair their serviceability, such as foreign matter or chips that affect the bore.

9. Packaging

9.1 For packaging, select from Specifications E 920, E 921, or Practice E 1157.

10. Keywords

10.1 disposable; pipet; plastic; serological

 **E 934 – 94 (2004)**

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).